TITAN EQUIPMENT

BUILT FOR LIFE'S CRITICAL WORK.

TITAN-COT-REF

LAYOUT EDITION 2

31 OCTOBER 2026

TITAN Mortuary Cots

CARE, INSPECTION & RECORDS SUPPLEMENT

NOT OPERATING INSTRUCTIONS



Model identification, inspection, care, service records and related TITAN families. This review edition does not replace original model-specific operating instructions or release the equipment for use.

Document reference: **TITAN-COT-REF** | Catalog: Mortuary Cots | Care and owner reference[1]

TITAN MORTUARY

1-888-792-9315 | service@mymortuarycooler.com

[1] Catalog identity and product reference:
https://www.mymortuarycooler.com/pages/titan-medical-mortuary-care-supplements

READ THIS BEFORE USING THE REFERENCE

Models: 500E / FS1-FC-500E, 1200XL / FS1-FC-1200XL, PRO S 650 / FS1-FC-650P, PRO XL / FS1-FC-1200P, 500T / FS1-FC-500T, P550 / FS1-FC-PT550

Identify the equipment first

Match the model, serial number, revision, fitted wheels, controls and order configuration to the actual equipment. Product names and similar appearance do not establish equal load limits, parts fit or operation.

Separate catalog facts from operating approval

Published catalog figures in this review are reference data, not engineering certification. Original assembly, lock/release, loading, adjustment, folding and service instructions still require model-specific verification. Keep this draft with, not instead of, the approved instructions.

Training and control

Only personnel trained for the specific equipment and task should use it. The facility must determine its staffing, transfer plan, infection-control procedure, route and emergency response. No universal one-person technique or staffing count is established here.

What this edition contains

Model sheets; preparation and inspection guidance; operating-document checkpoints; cleaning and storage guidance; service records; related TITAN families and accessory links. Church-truck finish and drape charts are included in the church-truck edition.

Release status

This edition is prepared for technical and owner review. Existing equipment-document release holds remain in place. It does not authorize an unverified load, accessory, vehicle restraint, slope, lubricant or repair method.

MODEL IDENTIFICATION & SPECIFICATIONS

500E / FS1-FC-500E

Published load capacity: 500 lb

Length × width: 76 × 22 inches

Minimum / maximum deck height: 14 / 32 inches

Maximum loading height: 24 inches

Wheels / net cot weight: 5 inches / 62 lb

Included restraints: Two 60-inch buckle straps



Model-specific boundary: Deck height and loading height are different measurements. Standard pouch and slider selections are only a starting point; verify the actual accessory dimensions and supplied configuration.

Catalog reference: https://www.mymortuarycooler.com/products/500e-standard-mortuary-cot

1200XL / FS1-FC-1200XL

Published load capacity: 1,000 lb

Length × width: 76 × 24.8 inches

Minimum height: 14 inches

Maximum loading height: 25 inches

Wheels / net cot weight: 6 inches / 110 lb

Included restraints: Three 60-inch securement straps



Model-specific boundary: 1200XL is the model name, not a 1,200 lb capacity. Confirm maximum deck height for the supplied revision. XL pouch, mattress and slider selections are not universal across all XL brands or platforms.

Catalog reference: https://www.mymortuarycooler.com/products/1000e-bariatric-mortuary-cot

MODEL IDENTIFICATION & SPECIFICATIONS

PRO S 650 / FS1-FC-650P

Published load capacity: 600 lb

Length × width: 77 × 22 inches

Minimum / maximum deck height: 14 / 32 inches

Maximum loading height: 24.5 inches

Wheels / net cot weight: 5 inches / 75 lb

Height adjustment: Five levels at each end



Model-specific boundary: 650 is the model designation; the current published load rating is 600 lb. Do not import another cot's vehicle-loading sequence or securement hardware.

Catalog reference: https://www.mymortuarycooler.com/products/650-pro-multi-level-mortuary-stretcher

PRO XL / FS1-FC-1200P

Published load capacity: 1,000 lb

Width: 24 inches

Minimum / maximum deck height: 14 / 33.25 inches

Loading height: 25 inches

Wheels / net cot weight: 6 inches / 97 lb

Leg construction: Stainless steel and aluminum alloy



Model-specific boundary: Overall length is not verified for the supplied revision. Do not substitute 1200XL dimensions or parts merely because both have a 1,000 lb published rating.

Catalog reference: https://www.mymortuarycooler.com/products/ame-pro-1000-1kp-oversized-mortuary-stretcher

MODEL IDENTIFICATION & SPECIFICATIONS

500T / FS1-FC-500T

Published load capacity: 600 lb

Length × width: 76.75 × 22 inches

High-position size: 76.75 × 22 × 32 inches

Low-position size: 76.75 × 22 × 11.8 inches

Vehicle loading height: 32 inches

Separately published deck height: 31 inches

Wheels / net cot weight: 5 inches / 63.9 lb



Model-specific boundary: The listing uses separate 32-inch high-position/loading figures and a 31-inch deck-height figure. Confirm measurement locations and actual vehicle compatibility; do not silently reconcile these figures.

Catalog reference: https://www.mymortuarycooler.com/products/pro-500t-high-loading-funeral-mortuary-stretcher

P550 / FS1-FC-PT550

Published numerical load rating: NOT VERIFIED: obtain before use

Maximum loading height: Up to 31 inches

Adjustable length × width: 63–81 × 22 inches

Lowest height: 20 inches

Highest head / foot height: 46 / 36 inches

Battery supply: Two 20V lithium batteries

Loading arrangement: Trackless; antler locking-device system



Model-specific boundary: P550 is not a verified 550 lb rating. Operating load, approved battery/charger details, emergency lowering and antler-system installation/release procedures must be documented before use. Do not treat a catalog description as vehicle securement approval.

Catalog reference: https://www.mymortuarycooler.com/products/titan-p550-power-first-call-cot-transport-stretcher

BEFORE-USE INSPECTION & READINESS

Identification and condition

Confirm the model label and applicable load rating are legible. Check frame members, joints, fasteners, contact surfaces and fitted accessories for missing parts, looseness, bends, cracks, corrosion or other damage.

Wheels and brakes

Check fitted wheels, caster attachment and braking components for damage, contamination and free movement using the actual model's approved inspection method. Brakes are positioning aids, not permission to leave an unstable or unattended load.

Support and lock condition

Check the actual model's locks, retaining devices, handles and support points. Do not use equipment whose lock engagement is uncertain or whose controls bind, drift or release unexpectedly. Never put hands into folding or scissor pinch points.

Restraints and vehicle system

Inspect straps, buckles, mattress attachment, rails if fitted and the model-specific vehicle locking system. A wheel brake is not a vehicle transport restraint. P550 battery, charger and emergency procedures require its original instructions.

Route and working area

Prepare a clear route and firm, level working area; identify thresholds, floor defects, door clearances and wet surfaces. No slope rating or all-terrain approval is established by this reference.

Stop-use rule

If any check fails, remove the equipment from service, label it clearly and contact support with the exact model and serial number. Do not attempt a loaded demonstration to prove a suspect mechanism is safe.

SETUP, MOVEMENT & DOCUMENT CHECKPOINTS

Before transfer

Identify the cot, published load limit, height controls, restraints and receiving surface. Keep standard and XL accessory sizes separate. Plan assistance and transfer control for the specific task.

Before vehicle loading

Compare the actual vehicle deck, required loading-height measurement and supplied locking-system installation. Confirm the approved loading, leg-support and securement procedure for that cot. Similar TITAN branding is not fit approval.

Before powered operation

For P550, obtain the documented load rating and approved battery, charging, emergency-lowering and antler-system procedures. No generic power-cot procedure is supplied by this reference.

Before transport or unloading

Check the actual vehicle restraint and cot securement under the approved procedure. Wheel brakes do not replace vehicle mounts. Confirm the complete unloading/leg-lock sequence before the cot is released from the vehicle.

Instructions still required

Original model-specific warnings, control identification, locking/unlocking sequence, load distribution limits, adjustment/loading/folding sequence, staffing requirements, maintenance schedule and authorized service procedures.

CLEANING, STORAGE & MAINTENANCE

After each use

Use cleaning and disinfection products approved for the actual frame, mattress, straps, buckles, electrical components and textile cover. Powered components and batteries need their own cleaning/isolation instructions; do not soak or pressure-wash electrical assemblies by assumption.

Disinfection

Follow the selected disinfectant label for dilution, application and contact time, and the equipment's verified compatibility instructions. Cleaning/disinfection is not a claim of sterilization. Follow the facility's exposure-control and PPE procedures.

Inspection records

Record date, model/serial, findings, corrective action and reviewer. Before-use checks do not replace the manufacturer's periodic service schedule. Obtain that schedule rather than inventing a monthly/annual interval.

Maintenance and repairs

For equipment, use only verified model-specific parts and approved repair procedures. For drapes, obtain textile-specific repair or replacement approval; do not add attachment holes or unapproved reinforcement. Do not modify the truck, bypass locks or substitute attachment hardware by assumption.

Storage and return to service

Store dry and protected in the approved stable configuration. Keep service access and exits clear. After repair, require documented authorized inspection and release; a cleaning entry alone is not a safe-use approval.

DISINFECTING & LUBRICATING

Prepare and separate the tasks

Follow facility exposure-control procedures and the required protective equipment. Remove the equipment from use for cleaning or service. Remove visible soil using the approved material-compatible cleaning method before disinfection. Decontaminate equipment before mechanical servicing; cleaning, disinfection and lubrication are different tasks.

Disinfect using an approved product

Confirm compatibility with every treated material and follow the disinfectant label for dilution, application, required wet contact time, rinsing where directed and drying. Do not mix chemicals or shorten contact time. Do not assume bleach, alcohol, iodine, phenolic products, steam or immersion is suitable for every TITAN item. Cleaning or disinfection is not a claim of sterilization.

Lubricate only verified service points

Use only the lubricant, amount, points and interval confirmed for the supplied model and revision. If that information is missing, obtain it from support before applying lubricant. Do not blanket-spray the frame or assume that locks, brake/friction surfaces, support pads, electrical parts or sealed bearings should be lubricated. Remove excess only by the approved method.

Apply the correct family guidance

Confirm the separate care of the frame, mattress, straps, buckles and cover. Follow the exact powered model's battery and electrical-isolation instructions before wet cleaning or service. Do not spray into controls, connectors, chargers or batteries. Use only confirmed service points; a cot model name does not establish a lubricant or interval.

Check and record before returning to service

Record the date, equipment identity, cleaning/disinfection product, label dilution and contact time, lubricant and approved points treated, findings, corrective work and service references. Check the applicable functions using the approved model instructions. Keep damaged equipment out of use until authorized repair and documented inspection; an entry in the log alone is not operating approval.

EXPLORE THE TITAN PRODUCT FAMILIES

TITAN Equipment has separate Cots and Tables lines. Links do not imply shared parts or procedures. Request approved instructions with exact model, serial/revision and fitted-control photos.

Equipment & support

Church trucks & drapes · Embalming & dressing tables · Mortuary cots & covers · TITAN product family · Technical support & document requests

Companion review manuals

Church Trucks · Embalming Tables · Folding Dressing Tables · Standard Folding Table · Cemetery Trucks · Frameless Drapes · Care & Inspection Records

Church-truck drape selection & care guide · Review-site copy.

TITAN DOCUMENTATION & SUPPORT

For model-specific instructions or service: 1-888-792-9315 · service@mymortuarycooler.com · support. Supply order, model, serial/revision and photos.

Rev A · October 30, 2026. Technical review only, not operating instructions, engineering approval or release. Existing warranty and terms apply.

EQUIPMENT & TRAINING RECORD

Facility / location: _____

Model / serial / revision: _____

Order / delivery date: _____

Original instruction number / revision: _____

Approved accessories / vehicle mount, if applicable: _____

Training record

Record the operator, trainer, exact instruction revision, demonstrated tasks and any restrictions. A signature here does not replace model-specific training.

Date	Operator / trainer	Instruction / task	Sign-off

INSPECTION & SERVICE LOG

INSPECTION RECORD · Follow the inspection and service intervals in the approved instructions for the exact model. Record actual work and retain supporting records.

Delivery date: _____ Model / serial: _____

Date: _____ Inspector / technician: _____
License / credential (if applicable): _____ Company: _____
Service / reference number: _____

Date: _____ Inspector / technician: _____
License / credential (if applicable): _____ Company: _____
Service / reference number: _____

Date: _____ Inspector / technician: _____
License / credential (if applicable): _____ Company: _____
Service / reference number: _____

Date: _____ Inspector / technician: _____
License / credential (if applicable): _____ Company: _____
Service / reference number: _____

Date: _____ Inspector / technician: _____
License / credential (if applicable): _____ Company: _____
Service / reference number: _____

Date: _____ Inspector / technician: _____
License / credential (if applicable): _____ Company: _____
Service / reference number: _____

WORK PERFORMED, FINDINGS & REFERENCE NOTES

Identify the inspection date; record findings, corrective work, parts and service / invoice references.

NOTES

TITAN EQUIPMENT

BUILT FOR LIFE'S CRITICAL WORK.

KEEP THIS DOCUMENT WITH THE EQUIPMENT

Care and records supplement. Not operating instructions. Keep the exact equipment instructions, training record and approved accessory information together.

PRODUCT DOCUMENTATION & SUPPORT

https://www.mymortuarycooler.com/pages/titan-medical-mortuary-care-supplements

To request the applicable operating instructions, provide an equipment-label photo, model and serial/lot. Do not send patient-identifying information.

TITAN EQUIPMENT / MORTUARY

Documentation and technical support
1-888-792-9315
service@mymortuarycooler.com
www.mymortuarycooler.com/pages/tech-support

DELIVER TO / EQUIPMENT LOCATION