TITAN EQUIPMENT

BUILT FOR LIFE'S CRITICAL WORK.

Standard Folding Table

Care & Owner Reference Manual

TITAN EQUIPMENT | TABLES



TECHNICAL REVIEW DRAFT

Model identification, inspection, care, service records and related TITAN families. This review edition does not replace original model-specific operating instructions or release the equipment for use.

Models: AME-MHFET / Multi-height folding

1-888-792-9315 | service@mymortuarycooler.com

TITAN-MHFET-REF • Rev A • October 30, 2026

Read this before using the reference

Identify the equipment first
Match the model, serial number, revision, fitted wheels, controls and order configuration to the actual equipment. Product names and similar appearance do not establish equal load limits, parts fit or operation.

Separate catalog facts from operating approval
Published catalog figures in this review are reference data, not engineering certification. Original assembly, lock/release, loading, adjustment, folding and service instructions still require model-specific verification. Keep this draft with, not instead of, the approved instructions.

Training and control
Only personnel trained for the specific equipment and task should use it. The facility must determine its staffing, transfer plan, infection-control procedure, route and emergency response. No universal one-person technique or staffing count is established here.

What this edition contains
Model sheets; preparation and inspection guidance; operating-document checkpoints; cleaning and storage guidance; service records; related TITAN families and accessory links. Church-truck finish and drape charts are included in the church-truck edition.

Release status
This edition is prepared for technical and owner review. Existing equipment-document release holds remain in place. It does not authorize an unverified load, accessory, vehicle restraint, slope, lubricant or repair method.

Model identification & specifications

AME-MHFET / Multi-height folding

Published load limit: 600 lb

Work surface: 80 × 30 inches

Height range: 32.5 to 38.5 inches; independent ends

Table weight: 100 lb

Top: 304 stainless steel trough-style

Drain: Center of foot end

Mobility / storage: Locking swivel casters; folding legs



Model-specific boundary: This is a manually adjusted folding platform, not the hydraulic platform. Obtain the exact release, support, locking and folding procedure; do not adjust a loaded table by assumption.

Catalog reference: https://www.mymortuarycooler.com/products/standard-embalming-table

Before-use inspection & readiness

Identification and condition

Confirm the model label and applicable load rating are legible. Check frame members, joints, fasteners, contact surfaces and fitted accessories for missing parts, looseness, bends, cracks, corrosion or other damage.

Wheels and brakes

Check fitted wheels, caster attachment and braking components for damage, contamination and free movement using the actual model's approved inspection method. Brakes are positioning aids, not permission to leave an unstable or unattended load.

Support and lock condition

Check the actual model's locks, retaining devices, handles and support points. Do not use equipment whose lock engagement is uncertain or whose controls bind, drift or release unexpectedly. Never put hands into folding or scissor pinch points.

Height system and drainage

Look for hydraulic leakage where fitted, damaged controls, unsupported table sections and obstructed drainage. Do not open a loaded hydraulic circuit or assume a foot-pedal function from another model.

Route and working area

Prepare a clear route and firm, level working area; identify thresholds, floor defects, door clearances and wet surfaces. No slope rating or all-terrain approval is established by this reference.

Stop-use rule

If any check fails, remove the equipment from service, label it clearly and contact support with the exact model and serial number. Do not attempt a loaded demonstration to prove a suspect mechanism is safe.

Setup, movement & document checkpoints

Before setup

Identify hydraulic, folding multi-height or fixed-frame multi-height construction. Confirm assembly, control functions, drain connection and the supplied-model locking procedure before use.

Before height adjustment

Use the exact model's adjustment and load-support procedure. Do not assign a pump or release function to an unidentified pedal or handle. Keep the under-table area clear.

Before preparation work

Prepare the drain and fluid collection arrangement using compatible facility equipment and infection-control procedures. A drain specification does not establish a building-plumbing connection or approve chemical exposure.

Before movement or service

Use the original loaded-movement restrictions; do not assume all casters permit loaded transport. Isolate from use before service and obtain the approved support/isolation method for the actual height mechanism.

Instructions still required

Original model-specific warnings, control identification, locking/unlocking sequence, load distribution limits, adjustment/loading/folding sequence, staffing requirements, maintenance schedule and authorized service procedures.

Cleaning, storage & maintenance

After each use

The AME-MHFET platform has a 304 stainless trough-style top and a folding tubular frame. Confirm cleaning and disinfectant compatibility with the top, frame, casters, locks and drain. Do not substitute another model's waxing, lubrication or maintenance products. This model's maintenance products and lubrication points are not verified.

Disinfection

Follow the selected disinfectant label for dilution, application and contact time, and the equipment's verified compatibility instructions. Cleaning/disinfection is not a claim of sterilization. Follow the facility's exposure-control and PPE procedures.

Inspection records

Record date, model/serial, findings, corrective action and reviewer. Before-use checks do not replace the manufacturer's periodic service schedule. Obtain that schedule rather than inventing a monthly/annual interval.

Maintenance and repairs

For equipment, use only verified model-specific parts and approved repair procedures. For drapes, obtain textile-specific repair or replacement approval; do not add attachment holes or unapproved reinforcement. Do not modify the truck, bypass locks or substitute attachment hardware by assumption.

Storage and return to service

Store dry and protected in the approved stable configuration. Keep service access and exits clear. After repair, require documented authorized inspection and release; a cleaning entry alone is not a safe-use approval.

Disinfecting & lubricating

Prepare and separate the tasks

Follow facility exposure-control procedures and the required protective equipment. Remove the equipment from use for cleaning or service. Remove visible soil using the approved material-compatible cleaning method before disinfection. Decontaminate equipment before mechanical servicing; cleaning, disinfection and lubrication are different tasks.

Disinfect using an approved product
Confirm compatibility with every treated material and follow the disinfectant label for dilution, application, required wet contact time, rinsing where directed and drying. Do not mix chemicals or shorten contact time. Do not assume bleach, alcohol, iodine, phenolic products, steam or immersion is suitable for every TITAN item. Cleaning or disinfection is not a claim of sterilization.

Lubricate only verified service points
Use only the lubricant, amount, points and interval confirmed for the supplied model and revision. If that information is missing, obtain it from support before applying lubricant. Do not blanket-spray the frame or assume that locks, brake/friction surfaces, support pads, electrical parts or sealed bearings should be lubricated. Remove excess only by the approved method.

Apply the correct family guidance
Treat the stainless trough, drain, frame, folding joints, casters and locking controls as distinct components. Use only the supplied model's confirmed lubricant map. A folding joint is not automatically an approved lubrication point; do not apply a hydraulic-table service procedure.

Check and record before returning to service
Record the date, equipment identity, cleaning/disinfection product, label dilution and contact time, lubricant and approved points treated, findings, corrective work and service references. Check the applicable functions using the approved model instructions. Keep damaged equipment out of use until authorized repair and documented inspection; an entry in the log alone is not operating approval.

Explore the TITAN product families

TITAN Equipment has separate Cots and Tables lines. Links do not imply shared parts or procedures. Request approved instructions with exact model, serial/revision and fitted-control photos.

Equipment & support
Church trucks & drapes · Embalming & dressing tables · Mortuary cots & covers · TITAN product family · Technical support & document requests

Companion review manuals
Church Trucks · Embalming Tables · Folding Dressing Tables · Mortuary Cots · Cemetery Trucks · Frameless Drapes

Church-truck drape selection & care guide · Review-site copy.

TITAN documentation & support

For model-specific instructions or service: 1-888-792-9315 · service@mymortuarycooler.com · support. Supply order, model, serial/revision and photos.

Rev A · October 30, 2026. Technical review only, not operating instructions, engineering approval or release. Existing warranty and terms apply.

Equipment & training record

Facility / location: _____

Model / serial / revision: _____

Order / delivery date: _____

Original instruction number / revision: _____

Approved accessories / vehicle mount, if applicable: _____

Training record

Record the operator, trainer, exact instruction revision, demonstrated tasks and any restrictions. A signature here does not replace model-specific training.

Date	Operator / trainer	Instruction / task	Sign-off

Inspection & service log

INSPECTION RECORD · Follow the inspection and service intervals in the approved instructions for the exact model. Record actual work and retain supporting records.

Delivery date: _____ Model / serial: _____

Date: _____ Inspector / technician: _____

License / credential (if applicable): _____ Company: _____

Service / reference number: _____

Date: _____ Inspector / technician: _____

License / credential (if applicable): _____ Company: _____

Service / reference number: _____

Date: _____ Inspector / technician: _____

License / credential (if applicable): _____ Company: _____

Service / reference number: _____

Date: _____ Inspector / technician: _____

License / credential (if applicable): _____ Company: _____

Service / reference number: _____

Date: _____ Inspector / technician: _____

License / credential (if applicable): _____ Company: _____

Service / reference number: _____

Date: _____ Inspector / technician: _____

License / credential (if applicable): _____ Company: _____

Service / reference number: _____

WORK PERFORMED, FINDINGS & REFERENCE NOTES

Identify the inspection date; record findings, corrective work, parts and service / invoice references.

Notes